

Transcatheter mitral valve replacement: there is no one-size-fits-all solution



Recambio mitral transcathéter: la talla única no existe

Juan F. Granada^{a,*} and Thomas Modine^b

^a Cardiovascular Research Foundation, Columbia University Medical Center, New York, NY, United States

^b Centre Hospitalier Universitaire de Lille, Institut Coeur Poumon, Heart Valve Clinic, Lille, France

Mitral regurgitation (MR) is the most prevalent form of valve disease in developed countries¹ and its prevalence increases with age affecting ~10% of people > 75 years.² MR is a very heterogeneous disease that damages not only the mitral valve apparatus but also its surrounding structures. In primary MR, mitral valve surgery is the therapy of choice in symptomatic patients or asymptomatic patients with left ventricular dysfunction. Surgical repair is generally preferred over replacement if technically feasible.^{3,4} In secondary MR, the surgical approach is still under discussion⁵ and it is spared for patients with indications for other surgical cardiac procedures (ie, coronary artery bypass graft).^{3,4}

Several studies suggest that a large proportion of MR patients are never treated due their high surgical risk⁶ and that such a conservative approach results in high re-hospitalization (~90%) and mortality rates (50%) within 5 years following the initial diagnosis.⁷ This population has become a clear unmet need and a target for the development of less invasive therapeutic approaches. Several transcatheter mitral valve repair (TMVR) technologies inspired by well-established surgical techniques have been developed and have already been approved for clinical use. The most commonly used, the MitraClip device (Abbott Vascular, United States) has reached more than 100 000 implants and demonstrated safety and efficacy in several MR subsets.⁸⁻¹⁰ Although clinical adoption continues to increase, edge-to-edge repair does not fully resolve MR and has some anatomical limitations too (ie, calcified leaflets) that prevent a wider use. Transcatheter mitral valve replacement (TMVR) offers a more universal concept for the management of mitral valve disease with a more predictable abolition of MR severity in a procedure that could be less invasive compared to current surgical techniques.¹¹

Important lessons have been learned from ongoing TMVR clinical studies. First, the patients screened for these trials, considered of high or prohibitive surgical risk display more complex anatomical substrates than originally thought that lead to very high rejection rates. Imaging sizing algorithms used to confirm patient eligibility are patient- and valve-design specific, but have not been standardized in all valve programs. The availability of different device sizes has also limited the wider adoption of this technology.

One of the biggest earliest technical concerns was the ability to achieve valve stability in the absence of sutures. To tackle this issue, several anchoring mechanisms have been developed resulting in high periprocedural success and low early

dislodgment rates.¹² The potential for left ventricular outflow tract obstruction is still the biggest Achilles' heel of this technology. Several factors including¹³ aorto-mitral-annular angle, degree of septal hypertrophy, the left ventricle size, and device protrusion into the cavity may contribute to left ventricular outflow tract obstruction. The short and mid-term valve leaflet performance has not been an issue to this day. Mean transvalvular gradients and paravalvular leaks have been similar to those obtained after surgical mitral valve replacement.

The rate of periprocedural complications varies based on the valve program we are dealing with. The mean 30-day all-cause mortality reported is ~13.6%.¹⁴ Approximately 4.6% of periprocedural death rate mainly due to unsuccessful TMVR deployment that ends up leading to conversion to open heart surgery. Also, issues with the management of access site are responsible for some deaths mainly associated with myocardial tears. The remaining deaths occur following the TMVR procedure. Transapical access has been associated with a higher rate of periprocedural complications (particularly bleeding) and mortality in TMVR procedures.¹⁴ The negative effects of thoracotomy in frail populations and the higher degree of myocardial injury associated with the transapical approach may be particularly deleterious in patients with reduced left ventricular ejection fraction.¹⁴ Finally, acute hemodynamic changes due to valve implantation in patients with severely depressed left ventricular ejection fraction (< 30%) is a very well-known phenomenon in the surgical field that worsens the prognosis of these patients.

Long-term data in a large cohort of patients is still lacking. In the largest series reported so far, no cases of structural valve degeneration, new occurrence of paravalvular leaks or valve dislocation requiring reintervention have been reported.¹⁴ However, a more systematic clinical and echocardiography follow-up of patients undergoing TMVR is crucial to provide consistent data on valve durability and structural valve failure in the future. Data on the risk of long-term thrombosis is scarce too. No episodes of clinically relevant valve thrombosis have been reported with other TMVR devices. Three-month anticoagulation therapy is currently recommended although no long-term data on the real thrombogenic profile of these devices has been reported yet.^{3,4,15}

TMVR is evolving to become a new alternative for treating patients with severe MR of very high or prohibitive surgical risk. The complexity of the mitral valve apparatus and the heterogeneity of

* Corresponding author: CRF-Skirball Center for Innovation, Cardiovascular Research Foundation, 8 Corporate Dr, Orangeburg, New York 10962, United States. E-mail address: jgranada@crf.org (J.F. Granada).

Online: 18-05-2020.

<https://doi.org/10.24875/RECICE.M20000122>

2604-7322/ © 2020 Sociedad Española de Cardiología. Published by Permanyer Publications. This is an open access journal under the CC BY-NC-ND 4.0 license.

the disease have limited the wide adoption of these technologies. Several devices are under clinical evaluation and the early experience gained with some of them proves the feasibility of their implantation. Larger studies including a larger number of patients are still needed to test the clinical performance of these technologies. Emerging transseptal TMVR systems have the potential to overcome some of the limitations of current transapical devices. However, technical and anatomical challenges will remain the same. The TMVR field is rapidly evolving, what is somehow clear now is that it will benefit a specific population subset and that there is no such thing as a one-size-fits-all solution.

CONFLICTS OF INTEREST

The CRF Skirball Center for Innovation has received research grant support from Edwards Lifesciences, Neovasc, Abbott Vascular, Sinomed and Cephea. J.F. Granada is one of the co-founders of Cephea.

REFERENCES

- Coffey S, Cairns BJ, Iung B. The modern epidemiology of heart valve disease. *Heart*. 2016;102:75-85.
- Nkomo VT, Gardin JM, Skelton TN, Gottdiener JS, Scott CG, Enriquez-Sarano M. Burden of valvular heart diseases: a population-based study. *Lancet*. 2006;368:1005-1011.
- Nishimura RA, Otto CM, Bonow RO, et al. 2017 AHA/ACC focused update of the 2014 aha/acc guideline for the management of patients with valvular heart disease: a report of the american college of cardiology/american heart association task force on clinical practice guidelines. *J Am Coll Cardiol*. 2017;70:252-89.
- Baumgartner H, Falk V, Bax JJ, et al. 2017 ESC/EACTS Guidelines for the management of valvular heart disease. *Eur Heart J*. 2017;38:2739-2791.
- Acker MA, Parides MK, Perrault LP et al. Mitral-valve repair versus replacement for severe ischemic mitral regurgitation. *N Engl J Med*. 2014; 370:23-32.
- Mirabel M, Iung B, Baron G, et al. What are the characteristics of patients with severe, symptomatic, mitral regurgitation who are denied surgery? *Eur Heart J*. 2007;28:1358-1365.
- Goel SS, Bajaj N, Aggarwal B et al. Prevalence and outcomes of unoperated patients with severe symptomatic mitral regurgitation and heart failure: comprehensive analysis to determine the potential role of MitraClip for this unmet need. *J Am Coll Cardiol*. 2014;63:185-186.
- Feldman T, Kar S, Elmariah S et al. Randomized Comparison of Percutaneous Repair and Surgery for Mitral Regurgitation: 5-Year Results of EVEREST II. *J Am Coll Cardiol*. 2015;66:2844-2854.
- Stone GW, Lindenfeld J, Abraham WT, et al. Transcatheter mitral-valve repair in patients with heart failure. *N Engl J Med*. 2018;379:2307-2318.
- Obadia JF, Messika-Zeitoun D, Leurent G, et al. Percutaneous repair or medical treatment for secondary mitral regurgitation. *N Engl J Med*. 2018; 379:2297-2306.
- Maisano F, Alfieri O, Banai S, et al. The future of transcatheter mitral valve interventions: competitive or complementary role of repair vs. replacement? *Eur Heart J*. 2015;36:1651-1659.
- Regueiro A, Granada JF, Dagenais F, Rodes-Cabau J. Transcatheter mitral valve replacement: insights from early clinical experience and future challenges. *J Am Coll Cardiol*. 2017;69:2175-2192.
- Meduri CU, Reardon MJ, Lim DS, et al. Novel multiphase assessment for predicting left ventricular outflow tract obstruction before transcatheter mitral valve replacement. *JACC Cardiovasc Interv*. 2019;12:2402-2412.
- Del Val D, Ferreira-Neto AN, Wintzer-Wehekind J, et al. Early experience with transcatheter mitral valve replacement: a systematic review. *J Am Heart Assoc*. 2019;8:e013332.
- Pagnesi M, Moroni F, Beneduce A, et al. Thrombotic risk and antithrombotic strategies after transcatheter mitral valve replacement. *JACC Cardiovasc Interv*. 2019;12:2388-2401.